

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 09/25/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08160

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:56
INTERNAL STANDARD	VERIFIED	03:56
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:57
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:58
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:59
BLANK TEST	0.00	03:59



SGT *Ralph J Bartlett*

9/25/08

SHT#160 # 162