

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/01/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08220

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:45
INTERNAL STANDARD	VERIFIED	04:45
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:46
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	04:47
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	04:48
BLANK TEST	0.00	04:49



565 *Ralph Bartlett* 10/1/08