

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 11/01/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08220

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:15
INTERNAL STANDARD	VERIFIED	02:15
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:16
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:17
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:18
BLANK TEST	0.00	02:19



SGS Supp / Bartlett 11/01/08 ST#620 #162