

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/03/2008

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08300

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:32
INTERNAL STANDARD	VERIFIED	23:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	23:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	23:34
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	23:35
BLANK TEST	0.00	23:36



SGS *Ralph Bartlett* 12/03/08