

# SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 12/10/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08300

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 02:09 |
| INTERNAL STANDARD       | VERIFIED | 02:09 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.09     | 02:10 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.09     | 02:11 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.09     | 02:12 |
| BLANK TEST              | 0.00     | 02:12 |



SGT Ralph J Bartlett

WCSO # 162

12-10-08