

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 12/24/2008

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08300

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:31
INTERNAL STANDARD	VERIFIED	01:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:32
REFERENCE STANDARD TEMP	33.9c	
REFERENCE STANDARD	0.09	01:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:34
BLANK TEST	0.00	01:35



SGT *Ralph Bartlett* #162

12/24/08