

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
State of New York Evidence Ticket

WARREN COUNTY SHERIFF'S OFFICE
BAC DATAMASTER 200066
FEBRUARY 06, 2008

OPERATOR NAME:
BARTLETT RALPH J
OPERATOR PERMIT NUMBER: 26222
OPERATOR AGENCY/DEPT:
WARREN COUNTY SHERIFF'S OFFICE
REFERENCE STANDARD NUMBER:
07300

--- SUPERVISOR MODE ---

BLANK TEST	.00	01:44
INTERNAL STANDARD	VERIFIED	01:44
REFERENCE STANDARD TEMP	33.9c	
REFERENCE STANDARD	.10	01:45
BLANK TEST	.00	01:45
REFERENCE STANDARD TEMP	33.9c	
REFERENCE STANDARD	.10	01:46
BLANK TEST	.00	01:46
REFERENCE STANDARD TEMP	33.9c	
REFERENCE STANDARD	.10	01:47
BLANK TEST	.00	01:48

I, R.J. BARTLETT, am employed by the

WARREN COUNTY SHERIFF'S OFFICE, a governmental agency, and, having been delegated by my employer to do so, certify that the above record of blood alcohol content analysis was made in the regular and ordinary course of business of this agency; that it is the regular and ordinary course of business of this agency to perform analyses of blood alcohol content and to make records of such analyses at the time they are performed; that the entries appearing on the above record were made at or soon after the time of the acts, transactions, occurrences or events stated thereon; that this record, if a copy, is a complete and exact duplicate of the original thereof; and that it is part of my employment responsibilities to maintain custody or control of this record.

FEB-06-08
DATE

Ralph J. Bartlett
SIGNATURE

PATROL SERGEANT WCSO #162
TITLE AND AGENCY