

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/24/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08300

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:28
INTERNAL STANDARD	VERIFIED	18:28
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:28
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:29
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	18:30
BLANK TEST	0.00	18:31



SGT Ruff/Budget #142

1/24/09