

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/03/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:36
INTERNAL STANDARD	VERIFIED	02:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:38
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:39



SGT *Ralph J Bartlett* # 162