

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/17/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:45
INTERNAL STANDARD	VERIFIED	00:45
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:45
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:46
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:47
BLANK TEST	0.00	00:48



505 Ralph Bartlett

SHERIFF # 162