

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/23/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:40
INTERNAL STANDARD	VERIFIED	03:40
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:41
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:43
BLANK TEST	0.00	03:44



SGT Ralph J Bartlett

WCSD

SHERIFF #162