

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/03/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 26222

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:02
INTERNAL STANDARD	VERIFIED	04:02
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:02
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:03
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:04
BLANK TEST	0.00	04:05



SGT RALPH J BARTLETT

WCSO

SHELD #162