

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/17/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:26
INTERNAL STANDARD	VERIFIED	04:26
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:26
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:27
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:28
BLANK TEST	0.00	04:29



SOS Ralph Bartlett

WCSO SH-162