

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 03/23/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:53
INTERNAL STANDARD	VERIFIED	04:53
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:53
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:54
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:55
BLANK TEST	0.00	04:56



SOS Ruffy Bartlett

WCSO #162