

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 04/01/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:13
INTERNAL STANDARD	VERIFIED	04:13
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:13
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:14
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:15
BLANK TEST	0.00	04:16



▼δ1W▼δ⊙†Ç◀◀◀

SOT *Ralph Bartlett*

WCSO

SHERID #162