

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 04/08/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:22
INTERNAL STANDARD	VERIFIED	17:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	17:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	17:23
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	17:24
BLANK TEST	0.00	17:25



SGT *Ralph Bartlett*

WCSO

SHIELD #162