

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 04/15/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:17
INTERNAL STANDARD	VERIFIED	03:17
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:17
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:19
BLANK TEST	0.00	03:20



SGS Rep of Results

WC SO

SHERCO # 102