

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106206  
 -----

Date: 04/23/2009

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:21
INTERNAL STANDARD	VERIFIED	05:21
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	05:21
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:23
BLANK TEST	0.00	05:24



SCS *Ralph Bartlett*

WCSO

SHERIFF #162

4/23/09