

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 04/29/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:01
INTERNAL STANDARD	VERIFIED	04:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:02
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:03
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:04
BLANK TEST	0.00	04:04



SG5 Ralph Bartlett WCSO #162