

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/07/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:59
INTERNAL STANDARD	VERIFIED	03:59
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	04:00
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	04:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	04:02
REFERENCE STANDARD TEMP	34.0c	
BLANK TEST	0.00	04:02



505 Sgt / Bartlett

WCSO #162