

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/13/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:00
INTERNAL STANDARD	VERIFIED	03:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:02
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:03
BLANK TEST	0.00	03:04



SGT Ralph Bartlett

WCSO #162