

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 05/20/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:34
INTERNAL STANDARD	VERIFIED	21:34
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	21:35
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	21:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	21:37
BLANK TEST	0.00	21:37



SGT Ralph J Bartlett

WCSO #162