

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 05/27/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:32
INTERNAL STANDARD	VERIFIED	01:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:34
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:35
BLANK TEST	0.00	01:36



SO Ralph Bartlett #162