

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/01/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:26
INTERNAL STANDARD	VERIFIED	02:26
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:27
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:28
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:29
BLANK TEST	0.00	02:30



SGT Ralph Bartlett Sheriff #162