

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/09/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09060

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:27 |
| INTERNAL STANDARD | VERIFIED | 00:27 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 00:27 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 00:28 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 00:29 |
| BLANK TEST | 0.00 | 00:30 |



SA Ralph Bartlett WCSO #162