

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 06/17/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:34
INTERNAL STANDARD	VERIFIED	21:34
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	21:34
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	21:35
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	21:36
BLANK TEST	0.00	21:37



SGT Ralph Bartlett 6/17/09