

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106206

Date: 07/02/2009

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:35
INTERNAL STANDARD	VERIFIED	03:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:39



SGR Ralph Bartlett WCSO #162