

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 07/08/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:22
INTERNAL STANDARD	VERIFIED	01:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:23
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:25



*See Page 1 of 2*

*WCSO #62 7/8/09*