

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 07/16/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:00
INTERNAL STANDARD	VERIFIED	04:00
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:02
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:03
BLANK TEST	0.00	04:03



Sgt Ralph Bartlett #162 7/16/09