

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 07/23/2009

Operator Name: KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:37
INTERNAL STANDARD	VERIFIED	07:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	07:38
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	07:39
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	07:40
BLANK TEST	0.00	07:40



Kenneth W Smith # 301

07/23/2009