

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/04/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:51
INTERNAL STANDARD	VERIFIED	03:51
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:51
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:52
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:53
BLANK TEST	0.00	03:54



SOT Rugsy Buzzer waso STCALD #162 8/4/09