

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 08/19/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:03
INTERNAL STANDARD	VERIFIED	04:03
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:03
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:04
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:05
REFERENCE STANDARD TEMP	34.0c	
BLANK TEST	0.00	04:06



505 Ralph Bartlett WSO #162 8/19/09