

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 09/02/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:45
INTERNAL STANDARD	VERIFIED	02:45
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:46
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:47
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:48
BLANK TEST	0.00	02:49



SGT Ralph Bartlett 9/24/09

WCSO #102