

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 09/10/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:23
INTERNAL STANDARD	VERIFIED	02:23
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:24
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:26
BLANK TEST	0.00	02:26



505 Ralph Bartlett Sheriff #102 9/10/09