

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 09/16/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:56
INTERNAL STANDARD	VERIFIED	02:56
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:57
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:58
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:59
BLANK TEST	0.00	02:59



SG Larry Burt WESO #162 9/16/09