

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 09/24/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:13
INTERNAL STANDARD	VERIFIED	03:13
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:14
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:15
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:16
REFERENCE STANDARD TEMP	34.0c	
BLANK TEST	0.00	03:17



SOB *Ralph Bartlett*

WCSO #162

9/24/09