

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 09/30/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:36
INTERNAL STANDARD	VERIFIED	03:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:39



*Sgt Ralph Bartlett*

SHERIFF # 162

09-30-09