

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/08/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:54
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:54
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56



55 Russy Russo 10/8/09