

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/17/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:07
INTERNAL STANDARD	VERIFIED	05:07
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:07
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:08
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:09
BLANK TEST	0.00	05:10



505 Rully Barrett

WCSO #162 10/17/09