

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 10/23/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:35
INTERNAL STANDARD	VERIFIED	04:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:38
BLANK TEST	0.00	04:39



SOS Ralph Bartlett

10/23/09

case # 162