

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 11/01/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:54
INTERNAL STANDARD	VERIFIED	04:54
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:55
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:56
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:57
REFERENCE STANDARD TEMP	34.0c	
BLANK TEST	0.00	04:58



SGS Ralph Bartlett #102 11/1/09