

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/03/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:47
INTERNAL STANDARD	VERIFIED	04:47
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:47
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:48
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:49
BLANK TEST	0.00	04:50



SOS *Ralph Bartlett*

WCSO SHERIFF #162 12/3/09