

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/09/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:20
INTERNAL STANDARD	VERIFIED	02:20
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:20
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:21
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:23



SGS *Ralph Bartlett* #162

12-09-09