

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 12/25/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:28
INTERNAL STANDARD	VERIFIED	20:28
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	20:29
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	20:30
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	20:31
BLANK TEST	0.00	20:31



SGT Ralph Bartlett 12/25/09 WERO #142