

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 01/01/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:48
INTERNAL STANDARD	VERIFIED	02:49
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:49
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:50
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:51
BLANK TEST	0.00	02:52



SGT Ralph Bartlett

WCSO # 162

01-01-10