

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/08/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:32
INTERNAL STANDARD	VERIFIED	09:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	09:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	09:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	09:34
BLANK TEST	0.00	09:35



SGT *Ralph Bartlett*

WCSO #162

1/8/10