

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 01/14/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:22
INTERNAL STANDARD	VERIFIED	05:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:23
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	05:24
BLANK TEST	0.00	05:25



SGT *Ralph Bartlett*

WARREN CO SHERIFF #162

1-14-10