

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106206

Date: 01/20/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:14
INTERNAL STANDARD	VERIFIED	04:14
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:14
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:15
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:16
BLANK TEST	0.00	04:17



SGT Ralph Bartlett weso #102

1/20/10