

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/02/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:50
INTERNAL STANDARD	VERIFIED	04:50
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:50
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:51
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:52
BLANK TEST	0.00	04:53



SCF Ralph J Bartlett 2-2-10 WPSO #162