

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106206

Date: 02/11/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:44
INTERNAL STANDARD	VERIFIED	02:44
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:44
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:45
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:47



565 *Ralph Bartlett* WOSO #162 02-11-10